

Date: 25th October, 2017

CORRIGENDUM-1

NIT No. : IUAC/NIT/09/FS/ 2017-18

Name of Tender: Supply of FTIR spectrometer with accessories

Please note the following corrections/ amendments in the above mentioned tender S. No
Published Should be read

[1]. **Interferometer** : Highly stable and permanently aligned

[Amended as: Highly stable permanently or dynamically aligned with 10 years warranty]

[2]. **Spectral range** : 6000 – 100 cm⁻¹ or better (Single source, single beam splitter and single detector without change of any optical components like source, beam splitter and detector)

[Amended as: 6000 – 100 cm⁻¹ or better (with dedicated beam splitters/detectors and source. However, these should be automatically switchable with computer software)

All other specifications and conditions will remain same.

A. O (S&P)
Inter University Accelerator Centre
Aruna Asaf Ali Marg
New Delhi- 110067